



An employee-owned company

August 7, 1998

11:39

CERTIFICATE OF ANALYSIS

WORKORDER:

9804395

SAMPLE SUMMARY

SENT **BUNCOMBE COUNTY SOLID WASTE**
TO: **3214 SPRING FOREST ROAD**
RALEIGH, N.C. 27616

ANALYZED **PBS&J Analytical Services**
BY: **6635 East Colonial Drive**
Orlando, Florida 32807

BOB HUNTER
919/876-6888X237 FAX 876-6840


Phone: (407) 277-4443
Fax: (407)382-8794

PROJECT: 28 080 04
PBS&J CONTACT: MÄY -
RECEIVED DATE: 04/24/98
REPORTED DATE: 05/21/98

WORK DESCRIPTION: **NEW FAC. S. WATERS**
TAKEN BY: **P. NERO**
TRANSPORTED: **GREYHOUND**
SAMPLE TYPES: **GRAB**
PO#:

State of Florida Certifications: E83011-Environmental, 83170-Drinking Water and Radiochemistry
CompQAP 860044G. State of North Carolina Certification: 547

SAMPLE DESCRIPTION	LAB ID	COLLECTED DATE/TIME
EQ BLK	01	04/22/98 09:00:00
SW-2	02	04/22/98 13:00:00



Tom French
Project Manager

Thank you for using PBS&J Analytical Services

CERTIFICATE OF ANALYSIS

RESULTS BY SAMPLE

SENT **BUNCOMBE COUNTY SOLID WASTE** ANALYZED
 TO: **3214 SPRING FOREST ROAD**
RALEIGH, N.C. 27616

PBS&J Analytical Services
 6635 East Colonial Drive
 Orlando, FL 32807

BOB HUNTER
919/876-6888X237 FAX 876-6840

Phone: (407) 277-4443
 Fax: (407) 382-8794

This is to certify that the following samples were analyzed using good laboratory practices to show the following results.

Sample ID: EQ BLK

Lab ID: 9804395-01D

Collected: 04/22/98 09:00:00

TEST	RESULT	UNITS	ANALYZED	BY
CONDUCTIVITY IN FIELD	0	umhos/cm		
pH IN FIELD	5.49	pH UNITS		
TEMPERATURE IN FIELD	10.7	DEG C		
DIBROMOCHLOROPROPANE	<0.02 U	ug/L	05/18/98	RP
ETHYLENE DIBROMIDE	<0.02 U	ug/L	05/18/98	RP
Appendix I Metals (TRACE)				
Antimony	<0.005 U	mg/L	04/28/98	MIM
Arsenic	<0.005 U	mg/L	04/28/98	MIM
Barium	<0.001 U	mg/L	04/28/98	MIM
Beryllium	<0.001 U	mg/L	04/28/98	MIM
Cadmium	<0.001 U	mg/L	04/28/98	MIM
Chromium	<0.010 U	mg/L	04/28/98	MIM
Cobalt	<0.005 U	mg/L	04/28/98	MIM
Copper	<0.005 U	mg/L	04/28/98	MIM
Lead	<0.004 U	mg/L	04/28/98	MIM
Nickel	<0.005 U	mg/L	04/28/98	MIM
Selenium	<0.005 U	mg/L	04/28/98	MIM
Silver	<0.005 U	mg/L	04/28/98	MIM
Thallium	<0.002 U	mg/L	04/28/98	MIM
Vanadium	<0.005 U	mg/L	04/28/98	MIM
Zinc	<0.020 U	mg/L	04/28/98	MIM
Appendix I Volatiles				
ACETONE	<20 U	ug/L	05/01/98	SI
ACRYLONITRILE	<8 U	ug/L	05/01/98	SI
BENZENE	<1 U	ug/L	05/01/98	SI
BROMOCHLOROMETHANE	<1 U	ug/L	05/01/98	SI
BROMODICHLOROMETHANE	<0.6 U	ug/L	05/01/98	SI
BROMOFORM	<1 U	ug/L	05/01/98	SI
CARBON DISULFIDE	<2 U	ug/L	05/01/98	SI
CARBON TETRACHLORIDE	<1 U	ug/L	05/01/98	SI
CHLOROETHANE	<1 U	ug/L	05/01/98	SI
CHLOROBENZENE	<1 U	ug/L	05/01/98	SI
CHLOROETHANE	<1 U	ug/L	05/01/98	SI
CHLOROFORM	<1 U	ug/L	05/01/98	SI
DIBROMOCHLOROMETHANE	<1 U	ug/L	05/01/98	SI
o-DICHLOROETHANE	<1 U	ug/L	05/01/98	SI
p-DICHLOROETHANE	<1 U	ug/L	05/01/98	SI
TRANS-1,4-DICHLORO-2-BUTE	<25 U	ug/L	05/01/98	SI
1,1-DICHLOROETHANE	<1 U	ug/L	05/01/98	SI
1,2-DICHLOROETHANE	<1 U	ug/L	05/01/98	SI
1,1-DICHLOROETHYLENE	<1 U	ug/L	05/01/98	SI
CIS-1,2-DICHLOROETHYLENE	<1 U	ug/L	05/01/98	SI
TRANS-1,2-DICHLOROETHYLENE	<1 U	ug/L	05/01/98	SI
1,2-DICHLOROPROPANE	<1 U	ug/L	05/01/98	SI
CIS-1,3-DICHLOROPROPENE	<1 U	ug/L	05/01/98	SI
TRANS-1,3-DICHLOROPROPENE	<1 U	ug/L	05/01/98	SI
ETHYLBENZENE	<1 U	ug/L	05/01/98	SI
2-HEXANONE	<10 U	ug/L	05/01/98	SI

CERTIFICATE OF ANALYSIS
RESULTS BY SAMPLE**Sample ID:** EQ BLK

Lab ID: 9804395-01C

Collected: 04/22/98 09:00:00

TEST	RESULT	UNITS	ANALYZED	BY
Appendix I Volatiles				
METHYL BROMIDE	<1 U	ug/L	05/01/98	SI
METHYL CHLORIDE	<1 U	ug/L	05/01/98	SI
METHYLENE BROMIDE	<1 U	ug/L	05/01/98	SI
METHYLENE CHLORIDE	<1 U	ug/L	05/01/98	SI
METHYL ETHYL KETONE	<20 U	ug/L	05/01/98	SI
METHYL IODIDE	<20 U	ug/L	05/01/98	SI
4-METHYL-2-PENTANONE	<10 U	ug/L	05/01/98	SI
STYRENE	<1 U	ug/L	05/01/98	SI
1,1,1,2-TETRACHLOROETHANE	<1 U	ug/L	05/01/98	SI
1,1,2,2-TETRACHLOROETHANE	<0.2 U	ug/L	05/01/98	SI
TETRACHLOROETHYLENE	<1 U	ug/L	05/01/98	SI
TOLUENE	<1 U	ug/L	05/01/98	SI
1,1,1-TRICHLOROETHANE	<1 U	ug/L	05/01/98	SI
1,1,2-TRICHLOROETHANE	<1 U	ug/L	05/01/98	SI
TRICHLOROETHYLENE	<1 U	ug/L	05/01/98	SI
TRICHLOROFLUOROMETHANE	<1 U	ug/L	05/01/98	SI
1,2,3-TRICHLOROPROPANE	<1 U	ug/L	05/01/98	SI
VINYL ACETATE	<10 U	ug/L	05/01/98	SI
VINYL CHLORIDE	<1 U	ug/L	05/01/98	SI
XYLENES	<1 U	ug/L	05/01/98	SI



An employee-owned company

August 7, 1998

11:39

CERTIFICATE OF ANALYSIS

WORKORDER:

SAMPLE SUMMARY

9804395

SENT **BUNCOMBE COUNTY SOLID WASTE**
TO: **3214 SPRING FOREST ROAD**
RALEIGH, N.C. 27616

ANALYZED BY: **PBS&J Analytical Services**
6635 East Colonial Drive
Orlando, Florida 32807

BOB HUNTER
919/876-6888X237 FAX 876-6840

Phone: (407) 277-4443
Fax: (407)382-8794

PROJECT: 28 080 04
PBS&J CONTACT: MÄY
RECEIVED DATE: 04/24/98
REPORTED DATE: 05/21/98

WORK DESCRIPTION: **NEW FAC. S. WATERS**
TAKEN BY: **P. NERO**
TRANSPORTED: **GREYHOUND**
SAMPLE TYPES: **GRAB**
PO#:

State of Florida Certifications: E83011-Environmental, 83170-Drinking Water and Radiochemistry
CompQAP 860044G. State of North Carolina Certification: 547

SAMPLE DESCRIPTION	LAB ID	COLLECTED DATE/TIME
EQ BLK	01	04/22/98 09:00:00
SW-2	02	04/22/98 13:00:00



Tom French
Project Manager

Thank you for using PBS&J Analytical Services

CERTIFICATE OF ANALYSIS

RESULTS BY SAMPLE

Sample ID: SW-2

Lab ID: 9804395-02D

Collected: 04/22/98

13:00:00

TEST	RESULT	UNITS	ANALYZED	BY
CONDUCTIVITY IN FIELD	64	umhos/cm		
pH IN FIELD	6.72	pH UNITS		
TEMPERATURE IN FIELD	93.0	DEG C		
DIBROMOCHLOROPROPANE	<0.02 U	ug/L	05/18/98	RP
ETHYLENE DIBROMIDE	<0.02 U	ug/L	05/18/98	RP
Appendix I Metals (TRACE)				
Antimony	<0.005 U	mg/L	04/28/98	MIM
Arsenic	<0.005 U	mg/L	04/28/98	MIM
Barium	0.086	mg/L	04/28/98	MIM
Beryllium	<0.001 U	mg/L	04/28/98	MIM
Cadmium	<0.001 U	mg/L	04/28/98	MIM
Chromium	<0.010 U	mg/L	04/28/98	MIM
Cobalt	<0.005 U	mg/L	04/28/98	MIM
Copper	<0.005 U	mg/L	04/28/98	MIM
Lead	<0.004 U	mg/L	04/28/98	MIM
Nickel	0.006 I	mg/L	04/28/98	MIM
Selenium	<0.005 U	mg/L	04/28/98	MIM
Silver	<0.005 U	mg/L	04/28/98	MIM
Thallium	<0.002 U	mg/L	04/28/98	MIM
Vanadium	0.006 I	mg/L	04/28/98	MIM
Zinc	0.02 I	mg/L	04/28/98	MIM
Appendix I Volatiles				
ACETONE	<20 U	ug/L	05/01/98	SI
ACRYLONITRILE	<8 U	ug/L	05/01/98	SI
BENZENE	<1 U	ug/L	05/01/98	SI
BROMOCHLOROMETHANE	<1 U	ug/L	05/01/98	SI
BROMODICHLOROMETHANE	<0.6 U	ug/L	05/01/98	SI
BROMOFORM	<1 U	ug/L	05/01/98	SI
CARBON DISULFIDE	<2 U	ug/L	05/01/98	SI
CARBON TETRACHLORIDE	<1 U	ug/L	05/01/98	SI
CHLOROETHANE	<1 U	ug/L	05/01/98	SI
CHLOROETHYLENE	<1 U	ug/L	05/01/98	SI
CHLOROFORM	<1 U	ug/L	05/01/98	SI
DIBROMOCHLOROMETHANE	<1 U	ug/L	05/01/98	SI
o-DICHLOROBENZENE	<1 U	ug/L	05/01/98	SI
p-DICHLOROBENZENE	<1 U	ug/L	05/01/98	SI
TRANS-1,4-DICHLORO-2-BUTE	<25 U	ug/L	05/01/98	SI
1,1-DICHLOROETHANE	<1 U	ug/L	05/01/98	SI
1,2-DICHLOROETHANE	<1 U	ug/L	05/01/98	SI
1,1-DICHLOROETHYLENE	<1 U	ug/L	05/01/98	SI
CIS-1,2-DICHLOROETHYLENE	<1 U	ug/L	05/01/98	SI
TRANS-1,2-DICHLOROETHYLENE	<1 U	ug/L	05/01/98	SI
1,2-DICHLOROPROPANE	<1 U	ug/L	05/01/98	SI
CIS-1,3-DICHLOROPROPENE	<1 U	ug/L	05/01/98	SI
TRANS-1,3-DICHLOROPROPENE	<1 U	ug/L	05/01/98	SI
ETHYLBENZENE	<1 U	ug/L	05/01/98	SI
2-HEXANONE	<10 U	ug/L	05/01/98	SI
METHYL BROMIDE	<1 U	ug/L	05/01/98	SI
METHYL CHLORIDE	<1 U	ug/L	05/01/98	SI
METHYLENE BROMIDE	<1 U	ug/L	05/01/98	SI
METHYLENE CHLORIDE	<1 U	ug/L	05/01/98	SI
METHYL ETHYL KETONE	<20 U	ug/L	05/01/98	SI
METHYL IODIDE	<20 U	ug/L	05/01/98	SI
4-METHYL-2-PENTANONE	<10 U	ug/L	05/01/98	SI
STYRENE	<1 U	ug/L	05/01/98	SI
1,1,1,2-TETRACHLOROETHANE	<1 U	ug/L	05/01/98	SI
1,1,2,2-TETRACHLOROETHANE	<0.2 U	ug/L	05/01/98	SI
TETRACHLOROETHYLENE	<1 U	ug/L	05/01/98	SI

August 7, 1998

11:41

CERTIFICATE OF ANALYSIS

RESULTS BY SAMPLE

Page 4

Sample ID: SW-2

Lab ID: 9804395-02C

Collected: 04/22/98 13:00:00

TEST	RESULT	UNITS	ANALYZED	BY
Appendix I Volatiles				
TOLUENE	<1 U	ug/L	05/01/98	SI
1,1,1-TRICHLOROETHANE	<1 U	ug/L	05/01/98	SI
1,1,2-TRICHLOROETHANE	<1 U	ug/L	05/01/98	SI
TRICHLOROETHYLENE	<1 U	ug/L	05/01/98	SI
TRICHLOROFLUOROMETHANE	<1 U	ug/L	05/01/98	SI
1,2,3-TRICHLOROPROPANE	<1 U	ug/L	05/01/98	SI
VINYL ACETATE	<10 U	ug/L	05/01/98	SI
VINYL CHLORIDE	<1 U	ug/L	05/01/98	SI
XYLENES	<1 U	ug/L	05/01/98	SI

**TABLE 12-2
DATA QUALIFIER CODES**

The following codes are used when reporting data values that either meet the specified description outlined below, or do not meet the quality control criteria of the laboratory:

Symbol:	Meaning:
A	Value reported is the mean (average) of two or more determinations. This code is to be used if the results of two or more discrete and separate samples are averaged. These samples shall have been processed and analyzed (e.g. laboratory replicate samples, field duplicates, etc.) independently. Do not use this code if the data are the result of replicate analysis on the same sample aliquot, extract or digestate. Under most conditions, replicate values shall be reported as individual analyses.
B	Results based upon colony counts outside the acceptable range. This code applies to microbiological tests and specifically to membrane filter colony counts. The code is to be used if the colony count is generated from a plate in which the total number of coliform colonies <u>exceeds</u> the method-indicated ideal ranges which are: Total Coliforms: 20 - 80 colonies Fecal Coliforms: 20 - 60 colonies/
J	Estimated value. This code shall be used in the following instances: <ol style="list-style-type: none">1. Surrogate recovery limits have been exceeded.2. No known quality control criteria exists for the component.3. The reported value failed to meet the established quality control criteria for either precision or accuracy.4. The sample matrix interfered with the ability to make any accurate determination.5. If the data is questionable because of improper laboratory or field protocols (e.g. composite sample was collected instead of a grab sample). <p>Note: a "J" value shall be accompanied by justification for its use. A "J" value shall not be used if any other code applies (e.g., K, L, M, T, V, Y, D). A "J" value <u>shall</u> be used if, in the analyst's opinion, the data results are <u>accurate</u>, and the QC failure does not affect the quality of the data results reported.</p>
K	Off-scale low. Actual value is known to be less than the value given. This code shall be used if: <ol style="list-style-type: none">1. The value is less than the lowest calibration standard <u>and</u> the calibration curve is known to be non-linear; or2. The value is known to be less than the reported value based on sample size, dilution, or some other variable. This code <u>shall not</u> be used to report values that are less than the laboratory practical quantitation limit or laboratory method detection limit.

TABLE 12-2 Continued
DATA QUALIFIER CODES

Symbol:	Meaning:
L	Off-scale high. Actual value is known to be greater than value given. To be used when the concentration of the analyte is above the acceptable level for quantitation (exceeds the linear range or highest calibration standard) and the calibration curve is known to exhibit a negative deflection.
M	When reporting chemical analyses: Presence of material is verified but not quantified; the actual value is less than the value given. The reported value shall be the laboratory Practical Quantitation Limit. This code shall be used if the level is too low to permit accurate quantification, but the estimated concentration is greater than the method detection limit. If the value is less than the method detection limit use "T" below. When reporting Oxygen Reduction Potential or Temperature: indicates a negative value.
N	Presumptive evidence of presence of material. This qualifier shall be used if: <ol style="list-style-type: none">1. The component has been tentatively identified based on mass spectral library search;2. There is an indication that the analyte is present, but quality control requirements for confirmation were not met (i.e. presence of analyte was not confirmed by alternate procedures).
O	Sampled, but analysis lost or not performed. Note: if reporting data to the U.S. Environmental Protection Agency Water Quality Storage and Retrieval (STORET) data base, a numerical value must be entered. Such values are not meaningful and shall not be used.
Q	Sample held beyond the accepted holding time. This code shall be used if the value is derived from a sample that was prepared or analyzed after the approved holding time restrictions for sample preparation or analysis.
T	Value reported is less than the laboratory method detection limit. The value is reported for informational purposes only; and shall not be used in statistical analysis.
U	Indicates that the compound was analyzed for but not detected. This shall be used to indicate that the specified component was not detected. The value associated with the qualifier shall be the laboratory method detection limit. Unless requested by the client, less than the method detection limit values shall not be reported (see "T" above).
V	Indicates that the analyte was detected in both the sample and the associated method blank. Note: the value in the blank shall not be subtracted from associated samples.

**TABLE 12-2 Continued
DATA QUALIFIER CODES**

Symbol:	Meaning:
Y	The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
Z	Too many colonies were present (TNTC), the numeric value represents the filtration volume.
I	The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit. It is reported as actual result.
?	Data is rejected and should not be used. Some or all of the following quality control data for the analyte was outside criteria, and the correct concentration of the analyte cannot be determined from the data, due to this QC failure: <ol style="list-style-type: none">1. Surrogate recovery limits have been exceeded.2. No known quality control criteria exists for the component.3. The reported value failed to meet the established quality control criteria for either precision or accuracy.4. The sample matrix interfered with the ability to make any accurate determination.5. If the data is questionable because of improper laboratory or field protocols (e.g. composite sample was collected instead of a grab sample). Note: a "?" value shall be accompanied by justification for its use. A "?" value shall be used if, in the analyst's opinion, the data results are <u>inaccurate</u> because of a QC failure.
*	Not analyzed due to interference.

If more than one code applies, only one code shall be reported. The code shall be selected based on the following hierarchy:

?
*, O
Y
V
B, K, L, M, I, U, T, Z
A, N, Q
J

**TABLE 12-2 Continued
DATA QUALIFIER CODES**

The following codes deal with certain aspects of field activities. The codes shall be used if the laboratory has knowledge of the specific sampling event. The codes shall be added by the organization collecting samples if they apply:

Symbol:	Meaning:
D	Measurement was made in the field (i.e. in-situ). This applies to any value (except pH, specific conductance, etc.) that was obtained under field conditions using approved analytical methods. Note: when data is to be entered into STORET, and the parameter code specifies a field measurement (e.g. "Field pH"), this code is not required.
E	Indicates that extra samples were taken at composite stations.
R	Significant rain in the past 48 hours. This code shall be used when the rainfall might contribute to a lower than normal value.
!	Data deviates from historically established concentration ranges.

FIELD LOGBOOK SAMPLING DATA SHEET

CLIENT: Buncombe County_
 PROJECT/SITE NAME: c/o PBS&J Ralieg_
 SITE ADDRESS: 3214 Spring Forest Rd._
 FIELD CONTACT NAME/PHONE: Ed Hilton, (919) 876-6888
 FIELD CONTACT ADDRESS: 3214 Spring Forest Rd._
 SAMPLING PERSONNEL: Vanessa May, Paul Nero

SAMPLING LOCATION: <i>SW-EQ B116</i>		SAMPLING DATE: <i>9/22/98</i>		SAMPLING TIME: <i>9:00</i>	
SAMPLE TYPE: <i>GRAB</i> / COMPOSITE		SAMPLE MATRIX: EQ B116 <i>EB</i>			
WELL DIAMETER	in.	PURGE RATE	gal/min		
TOTAL WELL DEPTH	ft.	PURGE START TIME			
DEPTH TO WATER	ft.	PURGE STOP TIME			
COLUMN HEIGHT	ft.	PURGING EQUIPMENT			
COLUMN VOLUME		SAMPLING EQUIPMENT / BAILER #			
MINIMUM PURGE VOLUME 3X OR 5X	Gallons	Column Volume (V) = 0.041 d ² h where: h = Column height (well depth-depth to water) d = diameter of well V = Volume of water			
ACTUAL PURGED VOLUME	Gallons	FIELD DECON PERFORMED PER PBS&J CQAP #910005?		YES / NO	
WELL CONSTRUCTION MATERIAL		DECON AND CALIBRATION WASTE CONTAINED?		YES / NO	
FUEL POWER UNIT UTILIZED AND LOCATED DOWNWIND FROM SAMPLING LOCATION?	YES / NO	PURGE WASTE CONTAINED?		YES / NO	
SAMPLE APPEARANCE			ENVIRONMENTAL CONDITIONS		
TINT	<i>clear</i>	AIR TEMPERATURE	<i>8.9</i>		
COLOR	<i>1</i>	WIND	<i>slight</i>		
TURBIDITY	<i>none</i>	RAIN	<i>light</i>		
ODOR	<i>1</i>	ATMOSPHERE	<i>cloudy</i>		
SHEEN	<i>1</i>	OTHER			
FIELD PARAMETERS					
TEMPERATURE (°C)	<i>10.7</i>	DISSOLVED OXYGEN (mg/L)	<i>8.84</i>		
CONDUCTIVITY (uMHOS/cm)	<i>0</i>	TURBIDITY (NTU)	<i>0</i>		
pH (UNITS)	<i>5.49</i>	SALINITY (ug/L)	<i>0</i>		
ADDITIONAL INFORMATION					
Chain of Custody #:					

FIELD LOGBOOK SAMPLING DATA SHEET

CLIENT: Buncombe County_
 PROJECT/SITE NAME: c/o PBS&J Ralieghe_
 SITE ADDRESS: 3214 Spring Forest Rd._
 FIELD CONTACT NAME/PHONE: Ed Hilton, (919) 876-6888
 FIELD CONTACT ADDRESS: 3214 Spring Forest Rd._
 SAMPLING PERSONNEL: Vanessa May, Paul Nero

SAMPLING LOCATION: <u>SW-1</u>		SAMPLING DATE: <u>4/22/98</u>	SAMPLING TIME: <u>13:00</u>
SAMPLE TYPE: <u>GRAB</u> / COMPOSITE		SAMPLE MATRIX: <u>SW</u>	
WELL DIAMETER	in.	PURGE RATE	gal/min
TOTAL WELL DEPTH	ft.	PURGE START TIME	
DEPTH TO WATER	ft.	PURGE STOP TIME	
COLUMN HEIGHT	ft.	PURGING EQUIPMENT	
COLUMN VOLUME		SAMPLING EQUIPMENT / BAILER #	
MINIMUM PURGE VOLUME 3X OR 5X	Gallons	Column Volume (V) = 0.041 d ² h where: h = Column height (well depth - depth to water) d = diameter of well V = Volume of water	
ACTUAL PURGED VOLUME	Gallons		
WELL CONSTRUCTION MATERIAL		FIELD DECON PERFORMED PER PBS&J CQAP #910005?	YES / NO
FUEL POWER UNIT UTILIZED AND LOCATED DOWNWIND FROM SAMPLING LOCATION?	YES / NO	DECON AND CALIBRATION WASTE CONTAINED?	YES / NO
		PURGE WASTE CONTAINED?	YES / NO
SAMPLE APPEARANCE		ENVIRONMENTAL CONDITIONS	
TINT	<u>clear light</u>	AIR TEMPERATURE	<u>8.7</u>
COLOR	<u>yellow</u>	WIND	<u>51.6kt</u>
TURBIDITY	<u>slight none</u>	RAIN	<u>light</u>
ODOR	<u>none</u>	ATMOSPHERE	<u>cloudy</u>
SHEEN	<u>none</u>	OTHER	
FIELD PARAMETERS			
TEMPERATURE (°C)	<u>13.0</u>	DISSOLVED OXYGEN (mg/L)	<u>8.47</u>
CONDUCTIVITY (uMHOS/cm)	<u>64</u>	TURBIDITY (NTU)	<u>40</u>
pH (UNITS)	<u>6.72</u>	SALINITY (ug/L)	<u>0</u>
ADDITIONAL INFORMATION			
Chain of Custody #:			



ENVIRONMENTAL LABORATORIES

6635 East Colonial Drive
Orlando, FL 32807
Phone: (407)277-4443 FAX: (407)382-8794

CHAIN OF CUSTODY RECORD AND ANALYSIS REQUEST

IN 0242

CLIENT NAME <i>Buncombe CNTy</i>			P.O. NUMBER			MATRIX AQUEOUS SOLID AIR OIL			<table border="1"> <tr><td>GENERALS</td><td>METAL</td><td>NUTRIENTS</td><td>VOA</td><td>SEMIVOLATILE</td><td>CYANIDE</td><td>PHENOL</td><td>BACTERIA</td><td>GREASE AND OIL</td><td>OTHER</td><td>OTHER</td></tr> <tr><td></td><td><i>API MUC</i></td><td></td><td><i>API VUC</i></td><td></td><td></td><td></td><td></td><td></td><td><i>603/0BCP</i></td><td></td></tr> </table>												GENERALS	METAL	NUTRIENTS	VOA	SEMIVOLATILE	CYANIDE	PHENOL	BACTERIA	GREASE AND OIL	OTHER	OTHER		<i>API MUC</i>		<i>API VUC</i>						<i>603/0BCP</i>	
GENERALS	METAL	NUTRIENTS	VOA	SEMIVOLATILE	CYANIDE																PHENOL	BACTERIA	GREASE AND OIL	OTHER	OTHER																	
	<i>API MUC</i>		<i>API VUC</i>																					<i>603/0BCP</i>																		
CLIENT ADDRESS			CLIENT PROJECT MANAGER <i>Ed Hilton</i>																																							
CITY, STATE, ZIP			PROJECT NUMBER																																							
SAMPLER(S) NAME(S) <i>P. Nero</i>			PROJECT NAME <i>Old Fac. S. Waters</i>			<input type="checkbox"/> STANDARD TAT <input type="checkbox"/> EXPEDITED TAT* * SUBJECT TO RUSH FEES REPORT DUE DATE: _____ REMARKS:																																				
STATION NO.	DATE	TIME	STATION/SITE IDENTIFICATION						NUMBER OF SAMPLE CONTAINERS SUBMITTED						LAB SAMPLE I.D.																											
	<i>4/22</i>	<i>9:00</i>	<i>EQ BIK</i>			<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>							<i>3</i>	<i>804395-01</i>																											
	<i>L</i>	<i>13:00</i>	<i>SW-1</i>			<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>							<i>3</i>	<i>02</i>																											
1. RELINQUISHED BY: (SIGNATURE) <i>[Signature]</i>			DATE <i>4/22/98</i>	TIME <i>17:00</i>	3. RELINQUISHED BY: (SIGNATURE) <i>[Signature]</i>			DATE <i>4/24/98</i>	TIME <i>11:00</i>	5. RELINQUISHED BY: (SIGNATURE)			DATE	TIME																												
2. RECEIVED BY: (SIGNATURE)			DATE	TIME	4. RECEIVED BY: (SIGNATURE)			DATE	TIME	6. RECEIVED BY: (SIGNATURE)			DATE	TIME																												

RECEIVED IN LABORATORY BY: (SIGNATURE) <i>[Signature]</i>		DATE <i>4/24/98</i>	TIME <i>11:00</i>	CONTAINERS INTACT? <input type="checkbox"/> YES <input type="checkbox"/> NO	CUSTODY SEALS INTACT? <input type="checkbox"/> YES <input type="checkbox"/> NO	SAMPLE CONTAINERS PREPARED BY:	DATE	TIME
LABORATORY REMARKS: <i>Rec @ 11</i>						SAMPLE CONTAINERS CHECKED BY:	DATE	TIME